
Midmark M11 - v1.0.0
Total Cycles: 2474

Name: RITTER 1
Sterilizer ID: V2178198
10/24/2022 07:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.7 F 0.0
2:00 75.0 F 0.0
4:00 78.7 F 0.0
6:00 106.9 F 1.9
8:00 141.6 F 4.7
10:00 170.4 F 7.7
12:00 193.2 F 11.3
14:00 207.2 F 2.7
16:00 217.3 F 6.5
18:00 228.0 F 10.3
20:00 234.7 F 8.6
22:00 238.0 F 10.7
24:00 234.8 F 7.6
26:00 247.8 F 14.1
28:00 258.6 F 20.1
30:00 268.7 F 26.5
32:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.9 F 29.3
2:00 272.0 F 28.6
3:00 272.3 F 29.3
4:00 272.5 F 29.0

Min 271.6 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 216.6 F 0.0
10:00 194.2 F 0.0
15:00 194.1 F 0.0
20:00 218.7 F 0.0
25:00 233.3 F 0.0
30:00 230.8 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:04
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:10:03
CYCLE COMPLETE

10/24/2022 08:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____