

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2114

Name: RITTER2  
Sterilizer ID: V2313228  
03/02/2023 10:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
2:00 177.6 F 3.3  
4:00 184.7 F 7.7  
6:00 207.0 F 13.0  
8:00 220.0 F 7.7  
10:00 234.6 F 13.3  
12:00 238.9 F 11.3  
14:00 238.8 F 10.4  
16:00 253.0 F 17.5  
18:00 265.2 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.8 F 29.8  
2:00 271.5 F 29.2  
3:00 272.7 F 29.9  
4:00 271.8 F 29.2  
  
Min 271.5 F 29.1  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.1 F 0.0  
5:00 221.2 F 0.0  
10:00 208.6 F 0.0  
15:00 209.0 F 0.0  
20:00 208.1 F 0.0  
25:00 217.0 F 0.0  
30:00 232.6 F 0.0  
Drying Complete

FILLING: 1:25  
HEATING: 19:45  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:23  
CYCLE COMPLETE

03/02/2023 11:23 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_