
Midmark M11 - v1.0.0
Total Cycles: 5007

Name: RITTER 1
Sterilizer ID: V2178198
10/17/2024 02:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.9 F 0.0
2:00 164.2 F 0.8
4:00 199.2 F 4.3
6:00 207.8 F 9.0
8:00 225.0 F 12.7
10:00 229.7 F 7.9
12:00 240.5 F 12.8
14:00 239.0 F 9.9
16:00 237.2 F 8.7
18:00 249.2 F 14.5
20:00 260.7 F 20.9
22:00 271.2 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 27.9
1:00 272.7 F 28.7
2:00 272.1 F 28.2
3:00 271.2 F 28.1
4:00 272.5 F 28.8

Min 271.1 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 231.4 F 0.0
10:00 231.2 F 0.0
15:00 232.3 F 0.0
20:00 227.6 F 0.0
25:00 229.1 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:45
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

10/17/2024 03:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____