
Midmark M11 - v1.0.0
Total Cycles: 5015

Name: RITTER 1
Sterilizer ID: V2178198
10/23/2024 07:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 79.1 F 0.0
4:00 95.6 F 0.0
6:00 118.9 F 0.1
8:00 145.0 F 0.4
10:00 171.9 F 0.9
12:00 193.5 F 3.5
14:00 208.0 F 6.3
16:00 217.7 F 9.7
18:00 228.9 F 13.5
20:00 230.1 F 8.0
22:00 240.7 F 12.7
24:00 238.0 F 9.8
26:00 238.3 F 8.9
28:00 248.2 F 14.2
30:00 259.6 F 20.4
32:00 270.0 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.4 F 28.6
2:00 271.7 F 28.3
3:00 272.4 F 28.6
4:00 271.5 F 28.1

Min 271.5 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 209.4 F 0.0
10:00 198.8 F 0.0
15:00 227.8 F 0.0
20:00 233.0 F 0.0
25:00 221.6 F 0.0
30:00 218.4 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 32:56
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:11:05
CYCLE COMPLETE

10/23/2024 09:09 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____