
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002451
11:32 24/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:08
PRESSURIZING 2:13
STERILIZING 3:12
136.5 C 222kPa 3:12
136.4 C 225kPa 3:42
136.4 C 225kPa 4:12
136.4 C 225kPa 4:42
136.1 C 223kPa 5:12
136.1 C 222kPa 5:42
136.2 C 223kPa 6:12
136.3 C 224kPa 6:42
136.3 C 224kPa 6:42
VENTING 6:42
AIR DRYING 7:33
CYCLE COMPLETE 32:05

Digital Signature #
62D3D40F68AD6ED7
