
Midmark M11 - v1.0.5
Total Cycles: 2316

Name: RITTER2
Sterilizer ID: V2313228
05/01/2023 09:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.6 F 0.0
2:00 151.0 F 0.5
4:00 162.3 F 3.3
6:00 183.6 F 7.6
8:00 205.0 F 12.4
10:00 220.9 F 6.9
12:00 232.7 F 12.0
14:00 234.3 F 9.2
16:00 239.1 F 9.9
18:00 246.3 F 14.2
20:00 258.6 F 21.1
22:00 269.7 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.5
1:00 272.6 F 30.9
2:00 272.6 F 30.3
3:00 272.4 F 30.8
4:00 272.6 F 30.5

Min 271.0 F 29.4
Max 273.3 F 31.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 212.4 F 0.0
10:00 229.1 F 0.0
15:00 231.6 F 0.0
20:00 229.8 F 0.0
25:00 229.9 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 22:59
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:13
CYCLE COMPLETE

05/01/2023 10:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____