
Midmark M11 - v1.0.0
Total Cycles: 2239

Name: RITTER 1
Sterilizer ID: V2178198
08/24/2022 12:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.9 F 0.0
2:00 144.4 F 0.0
4:00 160.7 F 0.4
6:00 189.0 F 2.2
8:00 209.2 F 6.5
10:00 226.7 F 11.2
12:00 227.1 F 6.6
14:00 239.9 F 12.0
16:00 237.8 F 9.6
18:00 237.7 F 8.9
20:00 249.5 F 14.9
22:00 261.6 F 21.6
24:00 272.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.0 F 28.3
2:00 272.5 F 28.7
3:00 272.1 F 28.8
4:00 271.8 F 28.2

Min 271.5 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 209.0 F 0.0
10:00 219.0 F 0.0
15:00 232.6 F 0.0
20:00 232.0 F 0.0
25:00 227.3 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:23
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:29
CYCLE COMPLETE

08/24/2022 01:04 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____