
Midmark M11 - v1.0.0
Total Cycles: 1446

Name: RITTER 1
Sterilizer ID: V2178198
12/09/2021 04:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.2 F 0.0
2:00 194.7 F 3.7
4:00 215.6 F 9.5
6:00 225.8 F 6.7
8:00 238.4 F 11.4
10:00 237.3 F 9.3
12:00 238.0 F 8.9
14:00 250.5 F 15.3
16:00 263.2 F 22.5
18:00 271.4 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.3 F 28.5
2:00 271.6 F 28.2
3:00 272.2 F 28.4
4:00 271.4 F 28.1

Min 271.3 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 236.8 F 0.0
10:00 228.5 F 0.0
15:00 230.3 F 0.0
Dry Cycle Aborted

FILLING: 3:56
HEATING: 18:03
STERILIZING: 4:00
VENTING: 3:06
DRYING: 19:31
TOTAL CYCLE: 00:48:36

12/09/2021 05:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____