
Midmark M11 - v1.0.5
Total Cycles: 2164

Name: RITTER2
Sterilizer ID: V2313228
03/22/2023 07:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.8 F 0.0
2:00 88.8 F 0.0
4:00 93.2 F 0.0
6:00 110.2 F 0.1
8:00 144.6 F 2.6
10:00 170.2 F 5.7
12:00 198.8 F 9.5
14:00 214.4 F 9.8
16:00 223.6 F 7.0
18:00 232.2 F 11.3
20:00 230.4 F 7.4
22:00 243.3 F 13.1
24:00 242.0 F 11.5
26:00 254.2 F 17.8
28:00 265.1 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.6 F 29.5
2:00 271.5 F 28.6
3:00 272.2 F 29.5
4:00 272.4 F 29.1

Min 271.1 F 28.5
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.8 F 0.0
5:00 226.4 F 0.0
10:00 214.4 F 0.0
15:00 215.9 F 0.0
20:00 217.7 F 0.0
25:00 223.1 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:54
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:07:51
CYCLE COMPLETE

03/22/2023 08:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____