
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001099
14:44 26/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:18
PRESSURIZING 4:19
STERILIZING 5:26
136.0 C 220kPa 5:26
136.4 C 223kPa 5:56
136.2 C 222kPa 6:26
136.4 C 222kPa 6:56
136.2 C 223kPa 7:26
136.4 C 225kPa 7:56
136.3 C 225kPa 8:26
136.4 C 225kPa 8:56
136.4 C 225kPa 8:56
VENTING 8:56
AIR DRYING 9:53
CYCLE COMPLETE 24:53

Digital Signature #
CCB9F5F026AC4142
