

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2614

Name: RITTER 1  
Sterilizer ID: V2178198  
11/28/2022 02:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.8 F 0.0  
2:00 204.8 F 5.0  
4:00 206.7 F 10.0  
6:00 222.4 F 6.2  
8:00 234.8 F 11.4  
10:00 234.9 F 8.8  
12:00 241.0 F 9.6  
14:00 247.3 F 14.2  
16:00 259.8 F 21.1  
18:00 271.2 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.5 F 29.4  
2:00 271.8 F 28.7  
3:00 272.8 F 29.6  
4:00 272.4 F 29.2  
  
Min 271.5 F 28.6  
Max 273.1 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.9 F 0.0  
5:00 237.6 F 0.0  
10:00 232.2 F 0.0  
15:00 226.5 F 0.0  
20:00 224.5 F 0.0  
25:00 227.1 F 0.0  
30:00 230.0 F 0.0  
Drying Complete

FILLING: 2:43  
HEATING: 18:42  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:38  
CYCLE COMPLETE

11/28/2022 03:25 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_