
Midmark M11 - v1.0.0
Total Cycles: 2910

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2023 03:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.9 F 0.0
2:00 189.6 F 4.4
4:00 195.5 F 9.2
6:00 218.2 F 8.4
8:00 231.3 F 9.2
10:00 237.7 F 9.8
12:00 241.9 F 12.0
14:00 240.0 F 10.5
16:00 253.4 F 17.2
18:00 264.7 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 29.1
2:00 271.3 F 28.5
3:00 272.5 F 29.2
4:00 271.6 F 28.5

Min 271.3 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 235.1 F 0.0
10:00 231.5 F 0.0
15:00 228.8 F 0.0
20:00 227.9 F 0.0
25:00 228.8 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:60
CYCLE COMPLETE

02/15/2023 04:57 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____