
Midmark M11 - v1.0.0
Total Cycles: 4139

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2023 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.1 F 0.0
2:00 170.2 F 3.0
4:00 176.7 F 6.9
6:00 205.1 F 11.2
8:00 217.7 F 5.4
10:00 226.4 F 10.6
12:00 230.7 F 7.0
14:00 243.6 F 13.0
16:00 242.7 F 11.6
18:00 255.5 F 18.2
20:00 266.8 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.5 F 29.0
2:00 271.5 F 28.3
3:00 272.0 F 28.9
4:00 272.4 F 28.8

Min 271.2 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.5 F 0.0
5:00 232.8 F 0.0
10:00 244.3 F 0.0
15:00 233.1 F 0.0
20:00 224.3 F 0.0
25:00 223.8 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 21:33
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:00:17
CYCLE COMPLETE

12/20/2023 12:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____