
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004744
16:50 21/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:56
STERILIZING 2:48
136.0 C 220kPa 2:48
136.0 C 221kPa 3:18
136.4 C 226kPa 3:48
136.3 C 222kPa 4:18
136.4 C 225kPa 4:48
136.3 C 223kPa 5:18
136.2 C 222kPa 5:48
136.4 C 225kPa 6:18
136.4 C 225kPa 6:18
VENTING 6:19
AIR DRYING 7:09
CYCLE COMPLETE 7:17

Digital Signature #
DDB7C0F66C011E00
