

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2306

Name: RITTER 1  
Sterilizer ID: V2178198  
09/09/2022 09:29 AM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.8 F 0.0  
2:00 160.6 F 1.0  
4:00 176.7 F 5.6  
6:00 201.2 F 10.8  
8:00 217.1 F 4.7  
10:00 234.1 F 11.1  
12:00 234.8 F 8.4  
14:00 239.6 F 9.6  
16:00 247.9 F 13.9  
18:00 260.7 F 21.1  
20:00 272.2 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 271.9 F 28.1  
2:00 272.6 F 28.6  
3:00 272.9 F 28.9  
  
Min 271.6 F 27.9  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 227.7 F 0.0  
5:00 223.8 F 0.0  
10:00 224.7 F 0.0  
Dry Cycle Aborted

FILLING: 1:04  
HEATING: 20:25  
STERILIZING: 3:00  
VENTING: 2:51  
DRYING: 11:49  
TOTAL CYCLE: 00:39:10

09/09/2022 10:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_