
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000176
14:49 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:19
PRESSURIZING 4:20
STERILIZING 5:46
136.5 C 223kPa 5:46
136.4 C 226kPa 6:17
136.3 C 226kPa 6:47
136.4 C 226kPa 7:17
136.3 C 225kPa 7:47
136.3 C 224kPa 8:17
136.4 C 227kPa 8:47
136.3 C 224kPa 9:17
136.3 C 224kPa 9:17
VENTING 9:17
AIR DRYING 10:21
CYCLE COMPLETE 25:21

Digital Signature #
DA389B469D9CD72A
