

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000395  
12:18 25/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:04  
PRESSURIZING 3:07  
STERILIZING 4:16  
136.4 C 232kPa 4:16  
136.4 C 233kPa 4:46  
136.4 C 232kPa 5:16  
136.3 C 233kPa 5:46  
136.0 C 230kPa 6:16  
136.2 C 230kPa 6:46  
136.4 C 233kPa 7:16  
136.4 C 233kPa 7:46  
136.4 C 233kPa 7:46  
VENTING 7:46  
AIR DRYING 8:39  
CYCLE COMPLETE 8:41  
-----

Digital Signature #  
8B0EC116B5AE1B11  
-----