
Midmark M11 - v1.0.0
Total Cycles: 4403

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2024 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.4 F 0.0
2:00 86.7 F 0.5
4:00 92.6 F 1.9
6:00 111.6 F 4.3
8:00 147.0 F 6.9
10:00 193.6 F 10.3
12:00 206.9 F 9.4
14:00 210.2 F 5.4
16:00 222.0 F 8.8
18:00 232.0 F 12.6
20:00 230.8 F 7.9
22:00 240.2 F 12.5
24:00 238.3 F 9.9
26:00 249.7 F 15.2
28:00 259.7 F 20.7
30:00 268.9 F 26.6
32:00 271.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.5 F 29.2
2:00 271.3 F 28.9
3:00 272.4 F 29.2
4:00 271.3 F 28.9

Min 271.2 F 28.7
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 232.7 F 0.0
10:00 229.0 F 0.0
15:00 234.0 F 0.0
20:00 223.5 F 0.0
25:00 221.5 F 0.0
30:00 220.3 F 0.0

Drying Complete

FILLING: 0:56
HEATING: 32:01
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:10:42
CYCLE COMPLETE

03/25/2024 04:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____