
Midmark M11 - v1.0.5
Total Cycles: 2903

Name: RITTER2
Sterilizer ID: V2313228
09/19/2023 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.7 F 0.0
2:00 70.0 F 0.0
4:00 77.4 F 0.0
6:00 107.6 F 0.0
8:00 149.3 F 0.5
10:00 176.1 F 2.0
12:00 199.5 F 5.3
14:00 209.1 F 8.7
16:00 220.5 F 12.2
18:00 224.6 F 6.0
20:00 234.8 F 10.1
22:00 236.0 F 9.7
24:00 242.5 F 12.1
26:00 239.9 F 10.6
28:00 252.0 F 16.5
30:00 262.5 F 22.6
32:00 272.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.0 F 28.8
2:00 272.5 F 29.3
3:00 271.4 F 28.8
4:00 272.0 F 28.8

Min 271.3 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.2 F 0.0
5:00 217.5 F 0.0
10:00 194.8 F 0.0
15:00 183.9 F 0.0
20:00 179.9 F 0.0
25:00 198.2 F 0.0
30:00 211.1 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 32:25
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:10:45
CYCLE COMPLETE

09/19/2023 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____