
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000333
8:43 08/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:23
STERILIZING 4:34
136.0 C 228kPa 4:34
136.4 C 233kPa 5:04
136.4 C 233kPa 5:34
136.3 C 230kPa 6:04
136.4 C 231kPa 6:34
136.2 C 230kPa 7:04
136.3 C 231kPa 7:34
136.4 C 232kPa 8:04
136.4 C 232kPa 8:04
VENTING 8:04
AIR DRYING 8:58
CYCLE COMPLETE 23:58

Digital Signature #
F3A041FDB37A2069
