
Midmark M11 - v1.0.0
Total Cycles: 3362

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2023 04:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.9 F 0.0
2:00 161.7 F 0.0
4:00 187.6 F 3.1
6:00 211.2 F 8.1
8:00 228.2 F 13.1
10:00 232.1 F 8.9
12:00 240.6 F 10.6
14:00 243.1 F 12.4
16:00 241.8 F 11.2
18:00 255.2 F 18.0
20:00 266.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.9
2:00 271.7 F 28.3
3:00 272.5 F 29.1
4:00 272.0 F 28.5

Min 271.5 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 207.1 F 0.0
10:00 216.9 F 0.0
15:00 220.6 F 0.0
20:00 228.1 F 0.0
25:00 227.7 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

06/08/2023 05:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____