
Midmark M11 - v1.0.0
Total Cycles: 2498

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2022 07:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.9 F 0.0
2:00 74.8 F 0.0
4:00 88.9 F 0.0
6:00 113.2 F 1.3
8:00 142.8 F 4.1
10:00 169.9 F 7.5
12:00 187.5 F 11.1
14:00 210.1 F 3.7
16:00 219.0 F 6.6
18:00 228.6 F 10.3
20:00 232.4 F 8.5
22:00 238.7 F 10.9
24:00 234.9 F 8.2
26:00 248.7 F 14.6
28:00 259.8 F 20.8
30:00 270.2 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 28.9
2:00 271.9 F 28.8
3:00 272.2 F 28.7
4:00 272.2 F 29.0

Min 271.5 F 28.2
Max 273.1 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.6 F 0.0
5:00 229.7 F 0.0
10:00 228.7 F 0.0
15:00 228.9 F 0.0
20:00 229.1 F 0.0
25:00 228.8 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:54
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:49
CYCLE COMPLETE

11/02/2022 09:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____