
Midmark M11 - v1.0.0
Total Cycles: 1307

Name: RITTER 1
Sterilizer ID: V2178198
10/27/2021 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.8 F 0.0
2:00 179.4 F 2.4
4:00 196.9 F 7.0
6:00 216.8 F 11.9
8:00 225.1 F 6.9
10:00 237.8 F 12.2
12:00 238.0 F 10.0
14:00 236.7 F 8.6
16:00 250.5 F 15.4
18:00 262.2 F 22.2
20:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.6 F 29.1
2:00 271.9 F 28.3
3:00 272.3 F 28.9
4:00 272.4 F 28.8

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 208.3 F 0.0
10:00 223.0 F 0.0
15:00 231.6 F 0.0
20:00 227.4 F 0.0
25:00 211.6 F 0.0
30:00 197.9 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 20:12
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

10/27/2021 11:10 AM
ACCEPTED: X
REJECTED:
Operator: U01