
Midmark M11 - v1.0.0
Total Cycles: 4866

Name: RITTER 1
Sterilizer ID: V2178198
09/04/2024 10:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.4 F 0.0
2:00 98.9 F 0.0
4:00 107.6 F 0.2
6:00 141.6 F 0.3
8:00 173.6 F 0.8
10:00 192.1 F 3.5
12:00 215.5 F 6.6
14:00 220.4 F 10.5
16:00 227.7 F 7.3
18:00 233.2 F 10.0
20:00 236.0 F 8.6
22:00 244.1 F 12.8
24:00 242.1 F 11.2
26:00 254.6 F 17.4
28:00 265.2 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.0
1:00 271.8 F 28.5
2:00 272.3 F 28.6
3:00 271.5 F 28.0
4:00 271.7 F 28.5

Min 271.2 F 27.9
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.4 F 0.0
5:00 221.9 F 0.0
10:00 207.3 F 0.0
15:00 219.0 F 0.0
20:00 233.8 F 0.0
25:00 237.4 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 29:51
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:08:18
CYCLE COMPLETE

09/04/2024 11:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____