
Midmark M11 - v1.0.0
Total Cycles: 1678

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2022 12:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.7 F 0.0
2:00 167.3 F 2.5
4:00 181.4 F 6.5
6:00 203.4 F 11.6
8:00 217.4 F 6.1
10:00 232.2 F 11.4
12:00 232.4 F 8.1
14:00 242.1 F 11.2
16:00 245.5 F 13.0
18:00 258.3 F 19.8
20:00 269.7 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.7 F 29.0
2:00 271.8 F 28.2
3:00 272.6 F 29.1
4:00 272.5 F 28.8

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 225.2 F 0.0
10:00 236.8 F 0.0
15:00 229.0 F 0.0
20:00 226.6 F 0.0
25:00 228.7 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 20:58
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:20
CYCLE COMPLETE

02/28/2022 01:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____