

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002143  
8:11 15/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:20  
PRESSURIZING 2:26  
STERILIZING 3:29  
136.4 C 224kPa 3:29  
136.3 C 224kPa 3:59  
136.2 C 222kPa 4:29  
136.4 C 224kPa 4:59  
136.4 C 224kPa 5:29  
136.0 C 222kPa 5:59  
136.2 C 221kPa 6:29  
136.4 C 224kPa 6:59  
136.4 C 224kPa 6:59  
VENTING 6:59  
AIR DRYING 7:48  
CYCLE COMPLETE 37:48  
-----

Digital Signature #  
B24EB5C88FEACB70  
-----