
Midmark M11 - v1.0.5
Total Cycles: 449

Name: RITTER2
Sterilizer ID: V2313228
10/14/2021 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.9 F 0.0
2:00 179.2 F 3.4
4:00 198.0 F 8.5
6:00 225.7 F 8.4
8:00 233.6 F 10.6
10:00 234.5 F 8.7
12:00 239.5 F 9.9
14:00 249.4 F 15.4
16:00 262.3 F 22.7
18:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.8
1:00 272.5 F 30.3
2:00 272.4 F 30.3
3:00 271.9 F 30.1
4:00 271.9 F 29.6

Min 270.4 F 28.7
Max 273.5 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 249.8 F 0.0
10:00 212.8 F 0.0
15:00 222.2 F 0.0
20:00 221.0 F 0.0
Dry Cycle Aborted

FILLING: 2:16
HEATING: 18:10
STERILIZING: 4:00
VENTING: 3:07
DRYING: 21:21
TOTAL CYCLE: 00:48:54

10/14/2021 09:42 AM
ACCEPTED: X
REJECTED:
Operator: U02