

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000741  
10:22 23/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:14  
PRESSURIZING 2:25  
STERILIZING 3:24  
136.0 C 226kPa 3:24  
136.2 C 232kPa 3:54  
136.2 C 232kPa 4:24  
136.3 C 230kPa 4:54  
136.4 C 233kPa 5:24  
136.4 C 231kPa 5:54  
136.4 C 233kPa 6:24  
136.3 C 233kPa 6:54  
136.3 C 233kPa 6:54  
VENTING 6:54  
AIR DRYING 7:46  
CYCLE COMPLETE 7:53  
-----

Digital Signature #  
423F82DF7D02CE44  
-----