
Midmark M11 - v1.0.5
Total Cycles: 3210

Name: RITTER2
Sterilizer ID: V2313228
12/13/2023 10:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 70.3 F 0.0
4:00 72.1 F 0.2
6:00 89.8 F 0.8
8:00 129.5 F 2.9
10:00 150.8 F 5.9
12:00 174.7 F 9.0
14:00 198.5 F 12.5
16:00 215.1 F 5.7
18:00 220.5 F 9.4
20:00 231.9 F 13.4
22:00 233.9 F 9.6
24:00 237.3 F 10.3
26:00 243.5 F 12.6
28:00 254.3 F 18.3
30:00 264.2 F 24.3
32:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.7
1:00 271.6 F 29.6
2:00 271.7 F 29.6
3:00 271.7 F 29.6
4:00 272.2 F 29.8

Min 271.3 F 29.5
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.2 F 0.0
5:00 224.8 F 0.0
10:00 227.7 F 0.0
15:00 222.1 F 0.0
20:00 215.8 F 0.0
25:00 213.5 F 0.0
30:00 212.8 F 0.0

Drying Complete

FILLING: 0:56
HEATING: 32:53
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:11:17
CYCLE COMPLETE

12/13/2023 11:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____