
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 001090
14:40 20/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:09
STERILIZING 4:25
136.0 C 221kPa 4:25
136.1 C 222kPa 4:55
136.4 C 226kPa 5:25
136.4 C 225kPa 5:55
136.4 C 222kPa 6:25
136.2 C 225kPa 6:55
136.4 C 225kPa 7:25
136.4 C 226kPa 7:55
136.4 C 226kPa 7:55
VENTING 7:55
AIR DRYING 8:49
CYCLE COMPLETE 23:49

Digital Signature #
626F44B1759D4854
