

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2230

Name: RITTER2  
Sterilizer ID: V2313228  
04/11/2023 11:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.2 F 0.0  
2:00 181.6 F 3.5  
4:00 204.8 F 9.0  
6:00 221.7 F 8.1  
8:00 233.7 F 10.7  
10:00 233.9 F 8.7  
12:00 238.8 F 10.0  
14:00 248.9 F 15.1  
16:00 261.9 F 22.5  
18:00 272.3 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 271.7 F 29.0  
2:00 271.8 F 28.9  
3:00 272.1 F 29.2  
4:00 272.4 F 29.5  
  
Min 271.5 F 28.9  
Max 272.5 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.4 F 0.0  
5:00 228.9 F 0.0  
10:00 224.3 F 0.0  
15:00 230.6 F 0.0  
20:00 231.8 F 0.0  
25:00 228.5 F 0.0  
30:00 228.7 F 0.0  
Drying Complete

FILLING: 2:11  
HEATING: 18:09  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:31  
CYCLE COMPLETE

04/11/2023 12:17 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_