
Midmark M11 - v1.0.5
Total Cycles: 3253

Name: RITTER2
Sterilizer ID: V2313228
12/27/2023 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.3 F 0.0
2:00 77.3 F 0.0
4:00 101.7 F 0.3
6:00 141.2 F 2.0
8:00 161.1 F 4.9
10:00 172.8 F 7.9
12:00 193.2 F 11.4
14:00 212.1 F 4.3
16:00 222.6 F 8.2
18:00 233.8 F 12.3
20:00 230.4 F 8.2
22:00 242.2 F 13.4
24:00 241.4 F 11.6
26:00 253.4 F 17.7
28:00 264.0 F 24.0
30:00 271.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.5 F 29.7
2:00 271.2 F 29.0
3:00 272.4 F 29.9
4:00 271.8 F 29.1

Min 271.1 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 228.4 F 0.0
10:00 207.8 F 0.0
15:00 203.0 F 0.0
20:00 202.2 F 0.0
25:00 201.9 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:08
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:13
CYCLE COMPLETE

12/27/2023 10:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____