

-----  
Midmark M11 - v1.0.5  
Total Cycles: 422

Name: RITTER2  
Sterilizer ID: V2313228  
10/06/2021 01:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.4 F 0.0  
2:00 149.3 F 2.2  
4:00 158.8 F 5.9  
6:00 185.5 F 10.8  
8:00 210.0 F 4.6  
10:00 223.6 F 10.2  
12:00 232.1 F 8.1  
14:00 241.1 F 12.4  
16:00 243.8 F 12.9  
18:00 255.9 F 19.6  
20:00 267.2 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.5  
1:00 272.4 F 29.7  
2:00 272.6 F 30.2  
3:00 271.8 F 29.6  
4:00 272.4 F 29.7  
  
Min 271.5 F 29.4  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.8 F 0.0  
5:00 222.9 F 0.0  
10:00 218.7 F 0.0  
15:00 212.6 F 0.0  
20:00 209.3 F 0.0  
25:00 190.0 F 0.0  
30:00 180.4 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 21:19  
STERILIZING: 4:00  
VENTING: 2:60  
DRYING: 30:00  
TOTAL CYCLE: 00:59:30  
CYCLE COMPLETE

10/06/2021 02:11 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01