
Midmark M11 - v1.0.0
Total Cycles: 5501

Name:
Sterilizer ID: V2178198
04/11/2025 03:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.4 F 0.0
2:00 108.0 F 0.9
4:00 138.2 F 3.1
6:00 143.5 F 6.0
8:00 164.7 F 9.1
10:00 181.5 F 12.3
12:00 206.5 F 4.2
14:00 212.3 F 7.9
16:00 223.9 F 11.8
18:00 228.7 F 6.9
20:00 238.7 F 11.7
22:00 235.6 F 8.9
24:00 246.7 F 14.5
26:00 257.0 F 20.0
28:00 266.9 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.3 F 29.9
2:00 272.6 F 29.6
3:00 271.5 F 29.1
4:00 271.3 F 29.4

Min 271.2 F 29.0
Max 273.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 184.9 F 0.0
10:00 176.9 F 0.0
15:00 175.9 F 0.0
20:00 185.7 F 0.0
25:00 192.1 F 0.0
30:00 201.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 29:40
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:08:31
CYCLE COMPLETE

04/11/2025 04:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____