
Midmark M11 - v1.0.5
Total Cycles: 2290

Name: RITTER2
Sterilizer ID: V2313228
04/24/2023 10:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.2 F 0.0
2:00 179.2 F 3.9
4:00 188.4 F 8.2
6:00 212.3 F 12.0
8:00 220.7 F 8.5
10:00 236.4 F 10.9
12:00 240.2 F 11.9
14:00 239.9 F 10.9
16:00 253.5 F 17.8
18:00 265.3 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.6 F 30.0
2:00 272.2 F 29.4
3:00 271.2 F 29.2
4:00 272.3 F 30.1

Min 271.2 F 29.0
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 217.9 F 0.0
10:00 245.4 F 0.0
15:00 226.5 F 0.0
20:00 221.7 F 0.0
25:00 221.2 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

04/24/2023 11:29 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____