
Midmark M11 - v1.0.5
Total Cycles: 3817

Name: RITTER2
Sterilizer ID: V2313228
07/26/2024 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.2 F 0.0
2:00 163.7 F 2.2
4:00 168.4 F 5.3
6:00 189.9 F 9.4
8:00 216.2 F 9.1
10:00 220.8 F 7.5
12:00 231.7 F 11.9
14:00 234.3 F 8.6
16:00 241.8 F 11.8
18:00 244.6 F 12.7
20:00 255.8 F 18.6
22:00 265.9 F 24.9
24:00 270.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.3 F 29.2
2:00 272.6 F 29.6
3:00 271.3 F 29.4
4:00 272.1 F 29.2

Min 271.2 F 29.0
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 231.5 F 0.0
10:00 226.1 F 0.0
15:00 228.0 F 0.0
20:00 227.3 F 0.0
25:00 227.1 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:41
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:32
CYCLE COMPLETE

07/26/2024 12:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____