
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001706
15:09 21/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:51
STERILIZING 4:02
136.4 C 225kPa 4:02
136.4 C 225kPa 4:32
136.4 C 225kPa 5:02
136.3 C 224kPa 5:32
136.4 C 225kPa 6:02
136.4 C 221kPa 6:32
136.1 C 223kPa 7:02
136.3 C 223kPa 7:32
136.3 C 223kPa 7:32
VENTING 7:32
AIR DRYING 8:20
CYCLE COMPLETE 8:23

Digital Signature #
F8690BF669505662
