
Midmark M11 - v1.0.0
Total Cycles: 2924

Name: RITTER 1
Sterilizer ID: V2178198
02/21/2023 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 217.9 F 0.0
2:00 207.0 F 4.7
4:00 215.5 F 10.3
6:00 226.6 F 6.2
8:00 237.1 F 12.5
10:00 238.8 F 10.6
12:00 237.4 F 9.5
14:00 252.7 F 16.8
16:00 265.3 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.1 F 28.9
2:00 271.5 F 28.7
3:00 272.2 F 28.9
4:00 271.4 F 28.5

Min 271.3 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 233.0 F 0.0
10:00 230.5 F 0.0
15:00 225.5 F 0.0
20:00 224.7 F 0.0
25:00 226.2 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 2:26
HEATING: 17:42
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:57:25
CYCLE COMPLETE

02/21/2023 01:45 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____