
Midmark M11 - v1.0.5
Total Cycles: 778

Name: RITTER2
Sterilizer ID: V2313228
02/10/2022 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.3 F 0.0
2:00 164.2 F 2.0
4:00 179.0 F 5.7
6:00 201.6 F 10.7
8:00 221.1 F 5.3
10:00 232.1 F 11.0
12:00 233.4 F 8.6
14:00 239.5 F 10.6
16:00 246.6 F 14.0
18:00 259.1 F 20.8
20:00 270.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.4
2:00 271.8 F 29.6
3:00 272.2 F 29.2
4:00 272.0 F 29.7

Min 271.2 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 233.5 F 0.0
10:00 241.6 F 0.0
15:00 227.1 F 0.0
20:00 217.9 F 0.0
25:00 217.5 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 20:53
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:57
CYCLE COMPLETE

02/10/2022 10:03 AM
ACCEPTED: X
REJECTED:
Operator: U02