
Midmark M11 - v1.0.0
Total Cycles: 2702

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2022 03:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.4 F 0.0
2:00 205.6 F 4.3
4:00 217.5 F 10.4
6:00 227.9 F 6.2
8:00 239.1 F 12.4
10:00 239.5 F 10.6
12:00 238.8 F 9.6
14:00 253.6 F 17.0
16:00 266.2 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.3 F 28.6
2:00 272.4 F 28.9
3:00 272.1 F 28.4
4:00 272.5 F 29.1

Min 271.6 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.2 F 0.0
5:00 221.8 F 0.0
10:00 232.5 F 0.0
15:00 233.3 F 0.0
20:00 226.2 F 0.0
25:00 222.0 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 17:31
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:04
CYCLE COMPLETE

12/20/2022 04:39 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____