
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001248
15:52 06/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:54
PRESSURIZING 3:56
STERILIZING 5:07
136.0 C 220kPa 5:07
136.4 C 225kPa 5:37
136.4 C 225kPa 6:07
136.1 C 222kPa 6:37
136.4 C 225kPa 7:07
136.4 C 224kPa 7:37
136.4 C 225kPa 8:07
136.4 C 225kPa 8:37
136.4 C 225kPa 8:37
VENTING 8:38
AIR DRYING 9:32
CYCLE COMPLETE 24:32

Digital Signature #
F7F37055BE0B20E2
