
Midmark M11 - v1.0.0
Total Cycles: 4477

Name: RITTER 1
Sterilizer ID: V2178198
04/22/2024 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.3 F 0.0
2:00 118.5 F 1.2
4:00 124.9 F 3.4
6:00 152.8 F 6.3
8:00 197.8 F 10.0
10:00 220.7 F 7.7
12:00 219.3 F 6.4
14:00 227.8 F 10.4
16:00 233.9 F 8.0
18:00 238.0 F 11.2
20:00 235.8 F 8.5
22:00 248.0 F 14.5
24:00 258.8 F 20.3
26:00 268.5 F 26.4
28:00 271.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.9 F 29.8
2:00 271.9 F 29.1
3:00 271.8 F 29.5
4:00 272.9 F 29.6

Min 271.2 F 28.9
Max 273.2 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 229.5 F 0.0
10:00 209.1 F 0.0
15:00 210.0 F 0.0
20:00 225.3 F 0.0
25:00 230.1 F 0.0
30:00 230.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:07
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:06:42
CYCLE COMPLETE

04/22/2024 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____