
Midmark M11 - v1.0.0
Total Cycles: 1499

Name: RITTER 1
Sterilizer ID: V2178198
01/05/2022 03:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.4 F 0.0
2:00 137.1 F 0.1
4:00 169.3 F 1.0
6:00 194.5 F 5.6
8:00 212.1 F 10.7
10:00 220.8 F 5.0
12:00 234.7 F 10.4
14:00 234.2 F 8.0
16:00 246.0 F 13.3
18:00 244.3 F 12.3
20:00 257.6 F 19.3
22:00 269.1 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.5
2:00 272.2 F 28.7
3:00 272.1 F 28.3
4:00 272.4 F 28.8

Min 271.4 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 169.7 F 0.0
5:00 201.6 F 0.0
10:00 213.7 F 0.0
15:00 215.5 F 0.0
20:00 206.8 F 0.0
25:00 202.1 F 0.0
30:00 198.7 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:02
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

01/05/2022 05:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____