
Midmark M11 - v1.0.0
Total Cycles: 4465

Name: RITTER 1
Sterilizer ID: V2178198
04/17/2024 09:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.4 F 0.0
2:00 167.5 F 1.7
4:00 173.4 F 4.8
6:00 191.5 F 9.0
8:00 208.2 F 13.4
10:00 224.7 F 6.4
12:00 232.4 F 10.9
14:00 230.8 F 7.1
16:00 243.1 F 12.9
18:00 242.0 F 11.1
20:00 253.5 F 17.0
22:00 264.2 F 23.2
24:00 271.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.1 F 28.6
2:00 272.7 F 28.9
3:00 271.4 F 28.6
4:00 272.0 F 28.6

Min 271.1 F 28.3
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 238.9 F 0.0
10:00 223.0 F 0.0
15:00 226.9 F 0.0
20:00 233.8 F 0.0
25:00 234.2 F 0.0
30:00 231.0 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 24:49
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:03:40
CYCLE COMPLETE

04/17/2024 10:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____