

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006556  
9:30 06/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:39  
PRESSURIZING 3:41  
STERILIZING 4:42  
136.0 C 220kPa 4:42  
136.4 C 223kPa 5:12  
136.3 C 220kPa 5:42  
136.1 C 221kPa 6:12  
136.3 C 221kPa 6:42  
136.4 C 223kPa 7:12  
136.3 C 219kPa 7:42  
136.4 C 222kPa 8:12  
136.4 C 222kPa 8:12  
VENTING 8:12  
AIR DRYING 9:04  
CYCLE COMPLETE 9:25

-----  
Digital Signature #  
3E583F95A27AC0F7  
-----