
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005314
9:02 04/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:06
STERILIZING 4:05
136.4 C 222kPa 4:05
136.3 C 222kPa 4:35
136.4 C 222kPa 5:05
136.4 C 224kPa 5:35
136.4 C 222kPa 6:05
136.4 C 222kPa 6:35
136.4 C 221kPa 7:05
136.4 C 222kPa 7:35
136.4 C 222kPa 7:35
VENTING 7:35
CYCLE ABORTED 7:36
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:36

Digital Signature #
C030F82ECE9EEBD3
