
Midmark M11 - v1.0.5
Total Cycles: 1650

Name: RITTER2
Sterilizer ID: V2313228
10/25/2022 08:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.9 F 0.0
2:00 170.1 F 2.4
4:00 176.0 F 6.3
6:00 209.1 F 11.6
8:00 221.5 F 7.2
10:00 235.5 F 12.7
12:00 235.3 F 9.9
14:00 236.2 F 8.7
16:00 247.9 F 15.0
18:00 259.7 F 21.7
20:00 270.9 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 271.9 F 30.2
2:00 272.3 F 30.0
3:00 272.7 F 30.5
4:00 271.8 F 29.9

Min 270.8 F 29.2
Max 273.3 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 214.3 F 0.0
10:00 231.9 F 0.0
15:00 230.4 F 0.0
20:00 224.5 F 0.0
25:00 222.2 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 20:39
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:50
CYCLE COMPLETE

10/25/2022 09:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____