
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002216
10:33 05/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:07
STERILIZING 4:15
136.4 C 225kPa 4:15
136.3 C 225kPa 4:45
136.3 C 224kPa 5:15
136.4 C 225kPa 5:45
136.3 C 224kPa 6:15
136.2 C 222kPa 6:45
136.4 C 225kPa 7:15
136.5 C 225kPa 7:45
136.5 C 225kPa 7:45
VENTING 7:45
AIR DRYING 8:39
CYCLE COMPLETE 9:25

Digital Signature #
0136ADA8AB41CCBD
