
Midmark M11 - v1.0.0
Total Cycles: 4050

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2023 11:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.8 F 0.0
2:00 93.9 F 0.3
4:00 108.5 F 1.8
6:00 141.2 F 4.4
8:00 165.4 F 7.6
10:00 187.7 F 11.7
12:00 204.9 F 3.2
14:00 215.8 F 6.5
16:00 226.2 F 10.3
18:00 231.5 F 8.6
20:00 238.1 F 10.8
22:00 236.3 F 8.2
24:00 248.0 F 14.4
26:00 259.0 F 20.4
28:00 269.3 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.9 F 29.3
2:00 272.1 F 28.6
3:00 271.7 F 28.7
4:00 272.9 F 29.3

Min 271.5 F 28.3
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 215.3 F 0.0
10:00 221.6 F 0.0
15:00 224.5 F 0.0
20:00 219.4 F 0.0
25:00 220.8 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:08
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:07:37
CYCLE COMPLETE

11/22/2023 12:39 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____