
Midmark M11 - v1.0.5
Total Cycles: 3063

Name: RITTER2
Sterilizer ID: V2313228
10/30/2023 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.4 F 0.0
2:00 156.4 F 2.0
4:00 182.0 F 6.0
6:00 203.8 F 10.7
8:00 218.9 F 5.1
10:00 228.1 F 10.2
12:00 234.3 F 8.2
14:00 242.6 F 12.2
16:00 245.6 F 13.3
18:00 257.9 F 20.0
20:00 269.0 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.0 F 29.3
2:00 272.7 F 29.6
3:00 272.5 F 29.9
4:00 271.4 F 29.4

Min 271.4 F 29.2
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.5 F 0.0
5:00 223.1 F 0.0
10:00 194.7 F 0.0
15:00 189.9 F 0.0
20:00 185.7 F 0.0
25:00 182.6 F 0.0
30:00 180.3 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:46
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

10/30/2023 12:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____